

# **BLANK PAGE**



IS: 4445 - 1967 (Reaffirmed 1977)

# Indian Standard SPECIFICATION FOR FILTER AND FILTER CHAMBER FOR BLOOD TRANSFUSION

(First Reprint MARCH 1986)

UDC 615.47:615.37



@ Copyright 1968

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

# Indian Standard

## SPECIFICATION FOR FILTER AND FILTER CHAMBER FOR BLOOD TRANSFUSION

Medical Glass Instruments and Appliances Sectional Committee, CPDC 12

Chairman

DR S. R. LELE

Members

SHRI J. K. WAD ( Alternate to Dr S. R. Lele )

COL A. K. BASU

SHRI N. C. DALAL

SHRI K. GANGAYYA

SHRI V. KRISHNAMOORTHY

SHRI PREM PRAKASH

SHRI SATVIR GUPTA ( Alternate )

SHRI B. M. L. SAXENA

MAJ SOHAN SINCH SHRI K. L. TALWAR

Shri A. N. Vig SHRI R. N. VIG ( Alternate )

SHRI A. B. RAO, 'Director (Consr Prod) Borosil Glass Works Ltd, Bombay

Directorate General of Armed Forces Medical Services (Ministry of Defence)

Scientific and Surgical Instrument Manufacturers & Traders Association, Bombay

Ministry of Health & Family Planning

Directorate General of Technical Development

Representing

National Physical Laboratory (CSIR), New Delhi

Defence Science Laboratory [Ministry of Defence (R & D)], New Delhi

Ministry of Defence (DGI)

Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

ISCOMED Industries Pvt Ltd, Bombay

Director General, ISI (Ex-officio Member)

Secretary Shri Vijai

Assistant Director (Consr Prod), ISI

Glass Instruments and Appliances Subcommittee, CPDC 12:2

Convener

COL A. K. BASU

Members

SHRI P. P. CHANDRA SHRI P. S. RAO ( Alternate )

Dr S. R. Lele

SHRI J. K. WAD ( Alternate )

DR S. C. NEOGY

SHRI B. M. L. SAKENA

SHRI K. L. TALWAR

DR S. A. TRIVEDI

Directorate General of Armed Forces Medical Services (Ministry of Defence)

Seraikella Glass Works (Private) Ltd, Singhbhum

Borosil Glass Works Ltd, Bombay

Scientific India Glass Co Ltd, Calcutta

Defence Science Laboratory [Ministry of Defence

(R & D)], New Delhi

Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

Research Laboratory Defence Defence (R & D)], Kanpur

[ Ministry of

INDIAN STANDARDS INSTITUTION

MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG

NEW DELHI

## Indian Standard

# SPECIFICATION FOR FILTER AND FILTER CHAMBER FOR BLOOD TRANSFUSION

#### O. FOREWORD

- 0.1 This Indian Standard was adopted by the Indian Standards Institution on 11 December 1967, after the draft finalized by the Medical Glass Instruments and Appliances Sectional Committee had been approved by the Consumer Products Division Council.
- **0.2** Preparation of standards on surgical instruments, medical equipment and apparatus including medical glass instruments has been taken up at the instance of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances (Government of India).
- **0.3** Great need has been felt for the standardization of glass instruments used in pathological work and blood transfusion. This standard covering the essential requirements of blood transfusion filter and filter chamber is expected to help in providing uniform equipment to all laboratories.
- 0.4 Assistance has been derived from B.S. 2463: 1962 'Transfusion equipment for medical use' issued by the British Standards Institution.
- 0.5 This standard is one of a series of Indian Standards on medical glass apparatus. Other specifications published so far in the series are:

IS: 3740-1966 Tubes, glass for pathological work

IS: 3741-1966 Tubes, sedimentation

IS: 3742-1966 Pipettes, dilution for haemocytometers

IS: 4067-1967 Tube, swab (West type), for throat

IS: 4068-1967 Ureometer, Doremus type

IS: 4069-1967 Urinometer

IS: 4087-1967 Pipette for haemoglobinometers and blood pipette for biochemical work

IS: 4363-1967 Drip counter

IS: 4364-1967 Pipettes, serological IS: 4444-1967 Bottles bacteriological

0.6 This standard contains clause 7.1 which calls for agreement between the purchaser and the supplier.

0.7 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS: 2-1960\*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

#### 1. SCOPE

1.1 This specification covers the requirements of filter and filter chamber used in the blood transfusion apparatus.

#### 2. MATERIALS

- 2.1 The filter tube and chamber shall be made from clear, colourless, neutral glass (for definition see IS: 1382-1961†). The glass shall pass the alkalinity test prescribed in IS: 2303-1963‡ for Type I glass.
- 2.2 The filter shall be made from silk.
- 2.3 The bung shall be made from rubber, conforming to IS: 3692-1965§.

#### 3. SHAPE AND DIMENSIONS

3.1 The shape and dimensions shall be as given in Fig. 1.

#### 4. WORKMANSHIP AND FINISH


- 4.1 The filter tube and chamber shall be well-annealed, free from bubbles and as far as possible, free from striae, stones and other visible defects (for definitions see IS: 1382-1961†). The ends shall be smoothly rounded in the flame. It shall be capable of being easily cleaned. It shall pass the thermal shock test, dry heat test and autoclave test specified in 5.1, 5.2 and 5.3 respectively.
- 4.2 The filter shall have a filtering area of not less than 32 cm<sup>2</sup>. The filter material shall be minimum of 80 percent as efficient as a sieve having a mesh with an average pore size of 0.212 mm square and a thread of 0.1 mm diameter (the reference filter material). The filter shall be disposed of after using it for one transfusion.

<sup>\*</sup>Rules for rounding off numerical values ( revised ).

<sup>†</sup>Glossary of terms relating to glass industry.

<sup>‡</sup>Method of grading glass for alkalinity.

<sup>§</sup>Specification for rubber closures (pharmaceutical).



All dimensions in millimetres ( Nom ).

Fig. 1 Filter and Filter Chamber for Blood Transfusion

#### 5. TESTS

- 5.1 Thermal Shock Test The filter tube and chamber shall be boiled in water for 30 minutes, then transferred to water at about 20°C. The glass shall not develop any chipping or cracking.
- 5.2 Dry Heat Test The filter tube and chamber shall be subjected to a dry heat test in a sterilizing oven at  $180^{\circ} \pm 2^{\circ}$ C for 30 minutes. The glass shall not show deterioration in any way nor develop any crack or chipping.
- 5.3 Autoclave Test The filter tube and chamber shall be autoclaved at a steam pressure of 1.4 kg/cm<sup>2</sup> for a period of 30 minutes. The glass shall not show deterioration in any way nor develop any crack or chipping.

#### 6. MARKING

- 6.1 The filter and chamber shall be marked with the name of the manufacturer, his initials or trade-mark.
- 6.1.1 The filter and chamber may also be marked with the ISI Certification Mark.

Note — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products as actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

#### 7. PACKING

- 7.1 The filter tube and chamber shall be packed as agreed to between the manufacturer and the purchaser. However, the recommended procedure is as follows:
  - 'Each set of filter tube and chamber shall be wrapped in suitable paper and packed in lots of 24 in suitable cartons'.



# INDIAN STANDARDS INSTITUTION

### Headquarters :

Headquarters :	
Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 1	10002
Telephones: 331 01 31, 331 13 75 Telegrams: 1	Manaksanstha
( Common	to all offices )
Regional Offices I	Telephone
*Western : Manakalaya, E9 MtDC, Marol, Andheri (East), BOMBAY 400093	6 32 92 95
†Eastern: 1/14 C. I. T. Scheme VII M, V. I. P. Road, Maniktola, CALCUTTA 700054	36 24 99
Southern : C. I. T. Campus, MADRAS 600113	41 24 42
Northern: B69 Phase VII, Industrial Focal Point, S. A. S. NAGAR 160051 (Punjab)	8 79 26
Branch Offices t	
'Pushpak' Nurmohamed Shaikh Marg, Khanpur,	12 63 48 2 63 49
'F' Block Unity Bldg, Narasimharaja Square, BANGALORE 560002	22 48 05
Gangotri Complex, Bhadbhada Road, T. T. Nagar, BHOPAL 462003	6 27 16
22E Kalpana Area, BHUBANESHWAR 751014	5 36 27
5-8-56C L. N. Gupta Marg, HYDERABAD 500001	22 10 83
R14 Yudhister Marg, C Scheme, JAIPUR 302005	6 98 32
117/418 B Sarvodaya Nagar, KANPUR 208005	4 72 92
Patliputra Industrial Estate, PATNA 800013	6 23 05
Hantex Bldg (2nd Floor), Rly Station Road, TRIVANDRUM 695001	32 27
Inspection Office ( With Sale Point ):	
Institution of Engineers (India) Building, 1332 Shivaji Nag PUNE 411005	ar, 5 24 35
*Sales Office in Bombay Is at Novelty Chambers, Grant Road	89 65 28
Street, Calcutta 700072  Street, Calcutta 700072	27 68 00